

SOP: Clinicians reporting adverse events

Version	Author(s)	Date	Changes	Approved by
1.0	Maria Skoog Anne Mette Plomgaard	29.05.12	Initial Version	Gorm Greisen
1.1	Maria Skoog Anne Mette Plomgaard	19.07.12	Minor changes	Gorm Greisen

CONTENT

1.0	Serious or suspected serious adverse events	1
2.0	Expected or unexpected non-serious adverse events	3

1.0 Serious or suspected serious adverse events

Events not considered related to the intervention are not to be reported!

Serious adverse events are to be assessed at 3, 6, 24, and 64 hours and at 7 days. (According to the protocol death and morbidity not related to the intervention are reported as outcome measurements)

1.1 How do I report a serious adverse reaction/serious device effect?

These must be reported to sponsor within 24hours. The events are reported in the eCRF, thereafter printed and faxed.

The procedure:

- Open the eCRF
- Choose: *Serious AE*
- Fill in the information required (see **Appendix A**)
- Press save
- Press the print-icon
- Print the Coverage: **Appendix B**
- Fax the printed Appendix B and the printed version of the eCRF to sponsor within 24 hours:

Professor Gorm Greisen
Neonataalklinikken
Rigshospitalet
Blegdamsvej 9
DK-2100
Denmark

Fax number: +45 3545 5025

- Call sponsor Gorm Greisen at **+45 3545 1326** or **+45 4046 0747** and inform about the incidence
- Save the original report for the GCP-monitor
- *Report the incidence to your local Ethic Committee according to national guidelines.*

Sponsor is according to protocol responsible for reporting the event to relevant national competent authorities.

What if the eCRF is unavailable?

- Print Appendix A or the SafeBoosC offline CRF
- Fill in the information required
- Print Appendix B, Coverpage and fax the information as described above

1.2 What is a serious adverse reaction/serious device effect?

An event that is related/suspected related to the intervention and:

- results in death
- is life-threatening
- requires prolongation of existing hospitalization
- results in persistent or significant disability, incapacity or requires intervention to prevent permanent impairment or damage

1.3 Which serious adverse reactions/serious device effects must be reported?

Device deficiencies: malfunction, misuse, or use error - if potential serious effect

SAE-NI: serious adverse event-near-incident

ESAR/ESADE: expected serious adverse reactions / expected serious adverse device effect

SUSAR/USADE: suspected unexpected serious adverse reaction / unanticipated serious adverse device effect

For definitions see *Appendix C*.

2.0 Expected or unexpected non-serious adverse events

Events not considered related to the intervention are not to be recorded!

Non-serious adverse events are to be assessed at 3, 6, 24, and 64 hours.

2.1 How do I report an expected or unexpected not-serious adverse reaction/not-serious device effect?

These events are reported in the eCRF (non-serious AE). The events are required data, and we recommend that this part of the eCRF is filled in at the end of the intervention period.

The procedure:

- Open the eCRF
- Choose (non-serious AE)
- Fill in the required information
- Press save

(Sponsor is according to protocol responsible for reporting the expected/unexpected not-serious events to the national competent authorities at the end of the trial.)

2.2 Which expected or unexpected non-serious adverse reactions/non-serious device effects are reported?

EAR/EADE expected adverse reactions / expected adverse device effect

UAR/UADE unexpected adverse reactions / unexpected adverse device effect

Device deficiencies: malfunction, misuse, or use error – not potentially serious effect.

For definitions see *Appendix C*.

Appendix A

eCRF - clinicians reporting serious adverse event

Patient ID: _____

Patient, date of birth: __ / __ / _____

Investigator/name of the person reporting the event:

Which event is being reported?

Only one can be chosen

Device deficiencies with potential serious effect
SAE-NI

Serious adverse event-near-incident

ESAR

Expected serious adverse reactions

ESADE

Expected serious adverse device effect

SUSAR

Suspected unexpected serious adverse reaction

USADE

Unanticipated serious adverse device effect

This report is:

Initiating report

Follow up report

Final report

Medical device

Which Medical Device was used:

Covidien, INVOS-oximeter, with Adult SomaSensor (SAFB-SM)

Hamamatsu, NIRO-200NX with Prope holder S type (A10963)

Was the event related to the Medical device – (NIRS-monitor)?

Yes

Possibly

No

Do not know

THE EVENT

When did you become aware of the event?

Date and time: __ / __ / __ __ __ __, __ : __

Has the event ended? Yes No

If yes, when did the event end?

Date and time: __ / __ / __ __ __ __, __ : __

Description of the event:

Category of the event:

More than one X is accepted. No X should be made if a SAE-IN is being reported.

Death

Life-threatening illness or injury

Prolongation of existing hospitalization

Permanent impairment of a body structure or body function

Necessity of medical or surgical intervention to avoid above mentioned

Was any action taken on the event? Yes No

If yes describe:

Event status:

Resolved

Resolved with sequelae

describe: _____

Ongoing

Death

Appendix B - Coverpage

**HASTER:
DENNE FAX SKAL VURDERES
INDEN 24 TIMER!**

SERIOUS ADVERSE EVENT SafeBoosC

Denne fax skal vurderes af Gorm Greisen indenfor 24 timer

**Gorm Greisen kan kontaktes på telefon 3545 1326 / 4046
0747**

**Hvis ikke Gorm Gerisen er tilgængelig, skal faxen leveres
til PhD-studerende Simon Hyttel-Sørensen eller Anne Mette
Plomgaard.**

Simon: 2812 4036

Anne Mette: 2858 8929

Appendix C

Definitions:

Adverse events (AE): any undesirable event occurring to a participant during a clinical trial, whether or not considered related to the trial intervention.

Serious adverse event (SAE): any adverse event that results in death, is life-threatening, requires prolongation of existing hospitalisation, result in persistent or significant disability or incapacity, or requires intervention to prevent permanent impairment or damage.

Adverse reactions (AR) / Adverse device effects (ADE): all untoward and unintended responses related to the interventions of application of the treatment guideline and/or cerebral oximeter.

Expected adverse reactions (EAR) / Expected adverse device effect (EADE): adverse reactions, we expect to be related to the interventions cerebral NIRS oximeter and/or the application of the treatment guideline.

Unexpected adverse reactions (UAR): All untoward and unintended responses, which are not, expected reactions, related to the interventions cerebral oximeter and/or application of the treatment guideline.

Serious adverse reactions (SAR): any adverse reaction that results in death, is life-threatening, requires prolongation of existing hospitalisation, results in persistent or significant disability or incapacity, or requires intervention to prevent permanent impairment or damage. This definition includes:

Serious adverse device effect (SADE): adverse device effect that has resulted in any of the consequences characteristic of a SAE, or are due to imprecise or incomplete results from diagnostic equipment e.g. incorrect or delayed diagnosis or incorrect or delayed treatment, and

Serious adverse effect-near-incidents (SAE-NI): device deficiencies that might have led to a SAE if a suitable action had not been taken, or intervention had not been made, or if circumstances had been less fortunate.

Expected serious adverse reactions (ESAR): any of the expected adverse reactions (listed above) that results in death, is life-threatening, requires prolongation of existing hospitalisation, results in persistent or significant disability or incapacity, or requires intervention to prevent permanent impairment or damage.

Suspected unexpected serious adverse reaction (SUSAR) / Unanticipated serious adverse device effect (USADE): an adverse reaction which is both serious and unexpected, i.e., not identified in the current risk analysis listed above (in expected adverse reactions). Included in this definition are the events with imminent risk of death, serious injury, or serious illness that requires prompt remedial action for other participants, users, or persons (including events that are of significant and unexpected nature such that they become alarming as a potential public health hazard or the possibility of multiple deaths occurring at short intervals) or a new finding to the event.

Device deficiencies: malfunction, misuse, or use error, e.g. rStO₂ values displayed despite displaced sensor, sensor positioned with light source facing away from skin surface, sudden malfunction of device that result in no or false rStO₂-values.